



CO-ORDINATION OF NOTIFIED BODIES
PPE Regulation 2016/425

PPE-R/00.043
Version 1

RECOMMENDATION FOR USE

Number of pages: 1		Approval stage :	Approved on :
Origin : Horizontal Committee		<input type="checkbox"/> Vertical Group	n/a
		<input checked="" type="checkbox"/> Horizontal Committee	30/05/2018
		<input type="checkbox"/> EU PPE Working Group	
Question related to	<input type="checkbox"/> PPE Regulation	<input checked="" type="checkbox"/> EN/prEN: 17025	<input type="checkbox"/> Other:
Article:	Annex:	Clause: 5.10.3.1 c)	
Key words: uncertainty of measurement			
Question: When notified bodies commission testing on test laboratories complying with EN/ISO/IEC 17025, and the reference specification includes pass / fail criteria, does the notified body have to make a specific request for uncertainty of measurement to be included in the test report?			
Solution: No. EN/ISO/IEC 17025 includes a clear requirement for uncertainties of measurement to be available and reported where the uncertainty might affect compliance with pass / fail criteria. In such cases, the test laboratory has to include the uncertainty.			